


Name of the Equipment	<b>Temperature and Humidity Test Chambers – 2 No's</b>
Purpose / Use	<ul style="list-style-type: none"> <li>• Used to carry out Constant &amp; Cyclic Humidity tests on subsystem</li> <li>• Confirming to MIL-STD-810F &amp; JSS 55555</li> </ul>
Photograph	
Key specification	<ul style="list-style-type: none"> <li>• Working Space : 1.0 x 1.0 x 1.0 M</li> <li>• Temperature Range : -40°C to 180°C</li> <li>• Relative Humidity : 10% to 98%</li> </ul>